



DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT

APPLICATION FOR HCAI PREAPPROVAL OF
MANUFACTURER'S CERTIFICATION (OPM)

OFFICE USE ONLY

APPLICATION #: OPM-0397

HCAI Preapproval of Manufacturer's Certification (OPM)

Type: ☐ New ☒ Renewal/Update

Manufacturer Information

Manufacturer: STERIS Corporation

Manufacturer's Technical Representative: Zach Miday

Mailing Address: 5960 Heisley Rd., Mentor, OH 44060

Telephone: (440) 280-2381

Email: Zachary_Miday@STERIS.com

Product Information

Product Name: V-PRO SERIES STERILIZERS

Product Type: Other Electrical & Mechanical Components

Product Model Number: V-Pro 1, V-Pro 1 Plus, V-Pro maX, V-Pro maX2 & VHP MD140X Sterilizer

General Description: Sterilizer used to sanitize medical instruments, gowns etc

Applicant Information

Applicant Company Name: EASE LLC.

Contact Person: Tiffany Tonn

Mailing Address: 1515 FAIRVIEW AVE, STE 205, MISSOULA, MT 59801

Telephone: (406) 541-3273

Email: tiffany@easeco.com

Title: Office Assistant

"A healthier California where all receive equitable, affordable, and quality health care"

STATE OF CALIFORNIA – HEALTH AND HUMAN SERVICES AGENCY





DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT

Registered Design Professional Preparing Engineering Recommendations

Company Name: EASE LLC

Name: Jonathan Roberson

California License Number: S4197

Mailing Address: 5877 Pine Ave., Suite 210, Chino Hills, CA 91709

Telephone: (951) 295-1892

Email: jon@EASECo.com

Certification Method

Testing in accordance with: ☐ ICC-ES AC156 ☐ FM 1950 ☐ ASHRAE 171 ☐ FEMA 461

☐ Other(s) (Please Specify):

*Use of criteria other than those adopted by the California Building Standards Code, 2025 (CBSC 2025) for component supports and attachments are not permitted. For distribution system, interior partition wall, and suspended ceiling seismic bracings, test criteria other than those adopted in the CBSC 2025 may be used when approved by HCAI prior to testing.

☒ Analysis

☐ Experience Data

☐ Combination of Testing, Analysis, and/or Experience Data (Please Specify):

HCAI Approval

Date: 8/8/2026

Name: William Staehlin

Title: Senior Structural Engineer

Condition of Approval (if applicable):

"A healthier California where all receive equitable, affordable, and quality health care"

STATE OF CALIFORNIA – HEALTH AND HUMAN SERVICES AGENCY





**EQUIPMENT ANCHORAGE
& SEISMIC ENGINEERING**

5877 Pine Ave, Ste. 210
Chino Hills, CA. 91709
Phn: (909) 606-7622

Office of Statewide Health Planning and Development
PREAPPROVAL OF MANUFACTURER'S CERTIFICATION
OPM-0397

THIS PREAPPROVAL CONFORMS TO THE 2019 CALIFORNIA BUILDING CODE

MANUFACTURER: **STERIS CORPORATION**
EQUIPMENT NAME: **V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS**

Sheet: 1 of 11

Date: 6/25/26

GENERAL NOTES

1. THIS HCAI PREAPPROVAL OF MANUFACTURER'S CERTIFICATION (OPM) IS BASED ON THE 2025 CBC. THE DEMANDS (DESIGN FORCES) FOR USE WITH THIS OPM SHALL BE BASED ON THE 2025 CBC
2. THIS DOCUMENT MAY ONLY BE USED WITH THE EXPRESS WRITTEN CONSENT OF THE MANUFACTURER LISTED ABOVE FOR THE SPECIFIC PROJECT SITE AND INSTALLATION LOCATION. THIS DOCUMENT IS INVALID WITHOUT SUCH CONSENT.
3. THIS PREAPPROVAL CONFORMS TO THE 2025 CALIFORNIA BUILDING CODE WHERE SDS IS NOT GREATER THAN 0.75, 1.90, 2.10 & 2.50.
4. FORCES PER ASCE 7-22 SECTION 13.3.1, EQUATIONS 13.3-1, 13.3-2 & 13.3-3,
WHERE $SDS = 1.90, I_p = 1.5, C_{AR} = 1.0, R_{po} = 1.5, z/h = 0, (R_u = 1.0, H_r = 1.0)$ AT CONCRETE SLAB AT OR BELOW GRADE. SEE FOLLOWING SHEETS FOR Ω_{∞}
WHERE $SDS = 2.10, I_p = 1.5, C_{AR} = 1.0, R_{po} = 1.5, z/h = 0, (R_u = 1.0, H_r = 1.0)$ AT CONCRETE SLAB AT OR BELOW GRADE. SEE FOLLOWING SHEETS FOR Ω_{∞}
WHERE $SDS = 0.75, I_p = 1.5, C_{AR} = 1.0, R_{po} = 1.5, Z/H = 0.50, (R_u = 1.3, H_r = 2.25)$ AT CONCRETE SLAB ON METAL DECK.
WHERE $SDS = 2.50, I_p = 1.5, C_{AR} = 1.0, R_{po} = 1.5, z/h = 0.95, (R_u = 1.3, H_r = 3.375)$ AT CONCRETE SLAB ON METAL DECK.
SEE FOLLOWING SHEETS FOR Ω_{∞}
5. THIS PREAPPROVAL COVERS ONLY THE SUPPORTS AND ATTACHMENTS OF THE EQUIPMENT TO THE STRUCTURE.
6. ALL DESIGN FORCES SHOWN ON THE DRAWINGS ARE FACTORED LOADS THAT SHALL BE USED FOR STRENGTH DESIGN.
7. CONCRETE SLAB ON METAL DECK DETAIL VALID FOR DEMANDS SHOWN AT ANY ELEVATION IN THE BUILDING. (i.e. $z/h = 0.50, 0.95$)
8. CONCRETE SLAB DETAIL VALID FOR DEMANDS SHOWN AT OR BELOW GRADE. (i.e. $z/h = 0$)
9. **RESPONSIBILITIES OF THE STRUCTURAL ENGINEER OF RECORD OF THE BUILDING**
 - A. PROVIDE SUPPORTING STRUCTURE TO SUPPORT WEIGHTS AND FORCES SHOWN IN ADDITION TO ALL OTHER LOADS.
 - B. VERIFY THAT THE INSTALLATION IS IN CONFORMANCE WITH THE 2025 CBC AND WITH THE DETAILS, MATERIAL AND GAGE OF THE UNIT WHERE ATTACHMENTS ARE MADE AGREE WITH THE INFORMATION SHOWN ON THE PREAPPROVAL DOCUMENTS.
 - C. VERIFY THAT PROJECT SPECIFIC SEISMIC PARAMETERS RESULT IN SEISMIC FORCES (E_h, E_v) THAT DO NOT EXCEED THE VALUES IN THIS OPM.
 - D. VERIFY THAT THE CONCRETE SLAB TO WHICH THE EQUIPMENT IS ANCHORED MEETS THE REQUIREMENTS OF THE APPLICABLE ICC ESR REPORT AND THIS OPM.
 - E. VERIFY THAT THE ANCHORS ARE AN ADEQUATE DISTANCE FROM ANY SLAB EDGES OR OPENINGS (SEE TYPICAL DETAIL ON SHEET 2).
 - F. VERIFY THAT ALL NEW OR EXISTING ANCHORS ARE AN ADEQUATE DISTANCE FROM THE UNIT ATTACHMENTS AND CHECK FOR INTERACTION WHERE OTHER ANCHORS ARE WITHIN 18" OR $6h_{ef}$ FROM THIS UNIT'S ANCHORS.



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26

SHEET

2

OF 11 SHEETS

10. EXPANSION ANCHORS:

A. ATTACHMENT IS TO BE MADE WITH THE ANCHORS LISTED BELOW AND INSTALLED AS DESCRIBED IN THE CORRESPONDING ICC REPORT.

Anchor Diameter	Concrete Type	Min. f _c (psi)	Anchor Type	ICC Report No.	Min. Embed.	Min. Spacing	Min. Edge Dist.	Min. Conc. Thickness	Torque Test	Direct Tension
3/8"	Sand Light Weight	3000	Hilti Kwik Bolt TZ2 (CARBON STEEL)	ESR-4266	2"	6.75"	12"	3.25" Over Flutes	30 FT-LB	N/A
3/8"	Sand Light Weight	3000	Hilti Kwik Bolt TZ2 (CARBON STEEL)	ESR-4266	2"	8"	16"	3.25"	30 FT-LB	1189 lb
3/8"	Normal Weight	3000	Hilti Kwik Bolt TZ2 (STAINLESS STEEL)	ESR-4266	2"	8"	14"	4"	30 FT-LB	1983 lb
3/8"	Normal Weight	3000	Hilti Kwik Bolt TZ2 (STAINLESS STEEL)	ESR-4266	2.50"	8"	18"	5"	30 FT-LB	2243 lb

B. THIS PREAPPROVAL ALLOWS FOR UP TO A MAXIMUM OF 2 ADJACENT CONCRETE SLAB EDGES, 14", 16" & 18" AWAY MINIMUM (i.e. - CORNER). SEE ADJACENT DETAIL FOR ADDITIONAL MINIMUM ALLOWABLE CONCRETE EDGE DISTANCES.

C. TESTING AND SPECIAL INSPECTION OF EXPANSION ANCHORS SHALL BE PERFORMED BY AN APPROVED INDEPENDENT AGENCY EMPLOYED BY THE FACILITY OWNER PER CBC 1704A & 1910A.5 AND CAC 7-149. ALL REPORTS SHALL BE SENT TO THE INSPECTOR OF RECORD, OWNER AND THE ARCHITECT OR ENGINEER IN RESPONSIBLE CHARGE.

- (i) DIRECT PULL TENSION TEST OR TORQUE TEST AT LEAST 50% OF THE ANCHORS.
- (ii) ACCEPTANCE CRITERIA:
 - DIRECT TENSION TEST: THE ANCHOR SHOULD HAVE NO OBSERVABLE MOVEMENT AT THE TEST LOAD. A PRACTICAL WAY TO DETERMINE OBSERVABLE MOVEMENT IS THAT THE WASHER BECOMES LOOSE.
 - TORQUE TEST: THE APPLICABLE TORQUE MUST BE ACHIEVED WITHIN THE FOLLOWING LIMITS: WEDGE TYPE : 1/2 TURN OF THE NUT


(iii) IF ANY ANCHOR FAILS, TEST ALL ANCHORS.

D. AVOID DAMAGING EXISTING STEEL REINFORCING IN CONCRETE SLAB WHEN INSTALLING CONCRETE EXPANSION ANCHORS.

E. PROVIDE FOR FULL THREAD ENGAGEMENT OF NUT & WASHER.

11. BOLTS THROUGH CONCRETE ON METAL DECK

- A. BOLTS SHALL BE TORQUED BY 3/4 TURN OF THE NUTS AFTER THE SNUG TIGHT (THE SNUG-TIGHT CONDITION IS DEFINED AS THE TIGHTNESS REQUIRED TO BRING THE CONNECTED PLIES INTO FIRM CONTACT) CONDITION IS ACHIEVED, UNLESS OTHERWISE NOTED.
- B. THROUGH BOLT HOLES SHALL BE 1/16" LARGER THAN BOLT SIZE (HOLE SIZE = BOLT SIZE + 1/16") FOR CONCRETE.
- C. THROUGH-BOLTS IN CONCRETE SHALL RECEIVE SPECIAL INSPECTION AND TESTING (THROUGH BOLTS WITH STEEL TO STEEL CONNECTION IN TENSION DO NOT REQUIRE TENSION TESTING) IN ACCORDANCE WITH REQUIREMENTS FOR POST-INSTALLED ANCHORS.



TYPICAL CONCRETE EDGE DETAIL



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2
& VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26


SHEET

3

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE SLAB AT OR BELOW GRADE



ANCHORS

MAX Sds	TYPE	DIAM	EFF EMBED	QTY	T _{SLAB}	*T _u (lb.)	*V _u (lb.)
190	HILTI KB TZ2	3/8"	2"	6	4"	1460	341
210	HILTI KB TZ2	3/8"	2.5"	6	5"	1628	377

* VALUES INCLUDE Ω_{op}

NOTES:

- FORCES ARE DETERMINED PER 2025 CALIFORNIA BUILDING CODE AND ASCE 7-22.
STRENGTH DESIGN IS USED. (EXAMPLE: $I_p=1.5$, $CAR=1.0$, $R_{po}=1.5$, $\Omega_{op}=1.5$, $R_U=1.0$, $H_f=1.0$, $z/h=0$)
- THIS PREAPPROVAL ENCOMPASSES WEIGHTS AND VERTICAL C.G. POSITIONS NOT EXCEEDING VALUES SHOWN.
- THIS PREAPPROVAL WAS PREPARED WITHOUT KNOWLEDGE OF ANY SITE CONDITION. COMPATIBILITY FOR USE WITH A SITE SHALL BE EVALUATED BY THE STRUCTURAL ENGINEER OF RECORD OF THE INSTALLATION (SEOR). USE REQUIRES APPROVAL BY THE SEOR.
- STRUCTURAL ENGINEER OF RECORD FOR THE INSTALLATION SHALL VERIFY ALL CONDITIONS, EVALUATE INTERACTION WITH ADJACENT EQUIPMENT AND ANCHORS, AND PROVIDE SUPPORT STRUCTURE DESIGNED TO SUPPORT WEIGHTS AND FORCES SHOWN IN COMBINATION WITH ALL OTHER LOADS THAT MAY BE PRESENT.
- SEE GENERAL NOTES: SHEETS 1 AND 2



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26

SHEET


4

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

$S_D \leq 1.90$

CONCRETE SLAB AT OR BELOW GRADE



DATE: 08/08/2026



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26

SHEET


5

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

$S_D \leq 2.10$

CONCRETE SLAB AT OR BELOW GRADE



DATE: 08/08/2026



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2
& VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26

SHEET


6

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE SLAB ON METAL DECK


(EXPANSION ANCHORS)



FRONT ELEVATION

NOTES:

- FORCES ARE DETERMINED PER 2025 CALIFORNIA BUILDING CODE AND ASCE 7-22.
STRENGTH DESIGN IS USED. (EXAMPLE: S_{ds}=0.75, I_p=1.5, C_{ar}=1.0, R_{po}=1.5, Ω_{ap}=1.5, R_μ=1.3, H_f=2.25, z/h=0.50)
HORIZONTAL FORCE (E_h) = 0.52 W_p
HORIZONTAL FORCE (E_{mh}) = 1.04 W_p (FOR CONCRETE ANCHORAGE)
VERTICAL FORCE (E_v) = 0.15 W_p
- THIS PREAPPROVAL ENCOMPASSES WEIGHTS AND VERTICAL C.G. POSITIONS NOT EXCEEDING VALUES SHOWN.
- THIS PREAPPROVAL WAS PREPARED WITHOUT KNOWLEDGE OF ANY SITE CONDITION. COMPATIBILITY FOR USE WITH A SITE SHALL BE EVALUATED BY THE STRUCTURAL ENGINEER OF RECORD OF THE INSTALLATION (SEOR). USE REQUIRES APPROVAL BY THE SEOR.
- STRUCTURAL ENGINEER OF RECORD FOR THE INSTALLATION SHALL VERIFY ALL CONDITIONS, EVALUATE INTERACTION WITH ADJACENT EQUIPMENT AND ANCHORS, AND PROVIDE SUPPORT STRUCTURE DESIGNED TO SUPPORT WEIGHTS AND FORCES SHOWN IN COMBINATION WITH ALL OTHER LOADS THAT MAY BE PRESENT.



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26

SHEET


7

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE SLAB ON METAL DECK

(Sds = 0.75, z/h = 0.50)



DATE: 08/08/2026

VIEWED FOR CODE COMPLIANCE
BY: William Staehlin
OPM-0397
CALIFORNIA BUILDING CODE, 2025



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26


SHEET

8

OF 11 SHEETS


SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE SLAB ON METAL DECK



NOTES:

- FORCES ARE DETERMINED PER 2025 CALIFORNIA BUILDING CODE AND ASCE 7-22.
STRENGTH DESIGN IS USED. (EXAMPLE: S_{ps}=2.50, I_p=1.5, C_{AR}=1.0, R_{po}=1.5, Ω_{op}=1.5, R_u=1.3, H_f=3.375, z/h=0.95)
HORIZONTAL FORCE (E_h) = 2.60 W_p
HORIZONTAL FORCE (E_{mh}) = 5.20 W_p (FOR CONCRETE ANCHORAGE)
VERTICAL FORCE (E_v) = 0.50 W_p
- THIS PREAPPROVAL ENCOMPASSES WEIGHTS AND VERTICAL C.G. POSITIONS NOT EXCEEDING VALUES SHOWN.
- THIS PREAPPROVAL WAS PREPARED WITHOUT KNOWLEDGE OF ANY SITE CONDITION. COMPATIBILITY FOR USE WITH A SITE SHALL BE EVALUATED BY THE STRUCTURAL ENGINEER OF RECORD OF THE INSTALLATION (SEOR). USE REQUIRES APPROVAL BY THE SEOR.
- STRUCTURAL ENGINEER OF RECORD FOR THE INSTALLATION SHALL VERIFY ALL CONDITIONS, EVALUATE INTERACTION WITH ADJACENT EQUIPMENT AND ANCHORS, AND PROVIDE SUPPORT STRUCTURE DESIGNED TO SUPPORT WEIGHTS AND FORCES SHOWN IN COMBINATION WITH ALL OTHER LOADS THAT MAY BE PRESENT.
- SEE GENERAL NOTES: SHEETS 1 AND 2



STERIS CORPORATION

DES. J. ROBERSON

SHEET

9

JOB NO. 14-2522


V-PRO 1, 1 PLUS, maX, maX 2
& VHP MD140X STERILIZERS

DATE 6/25/26

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE SLAB ON METAL DECK



DATE: 08/08/2026



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26


SHEET

10

OF 11 SHEETS

SEISMIC SUPPORTS & ATTACHMENTS

CONCRETE DETAILS



STERIS CORPORATION

V-PRO 1, 1 PLUS, maX, maX 2 & VHP MD140X STERILIZERS

DES. J. ROBERSON

JOB NO. 14-2522

DATE 6/25/26


SHEET

11

OF 11 SHEETS


SEISMIC SUPPORTS & ATTACHMENTS

BRACKET DETAILS




PLAN

BRACKET MATERIAL
5/16" THK, ASTM A1011, Fy=30 KSI MIN



FRONT



SIDE

BRACKET DETAIL (B)

