

Office of Statewide Health Planning and Development



APPLICATION FOR PREAPPROVAL

SPECIAL SEISMIC CERTIFICATION OF EQUIPMENT AND COMPONENTS

	For Office Use Only	
	APPLICATION NO. Chec OSP - 0281-10	k whether application is: NEW X RENEWAL
1.0	TOSHIBA MEDICAL SYSTEMS	Greg Patterson
	Manufacturer	Manufacturer's Technical Representative
	2441 Mich	elle Drive, Tustin, CA 92681
		Mailing Address
	(714) 669-7842	
	Telephone	E-mail Address
2.0	KALARE	R/F MEDICAL IMAGING
	Product Name	Product Type
	SEE AT	TTACHMENT 1
		roduct identification numbers and/or serial numbers)
	Troduct Model No (Elst all unique p	roduct identification numbers and/or serial numbersy
	General Description: A multi-component radi system.	ography/fluoroscopy (R/F) medical diagnostic imaging
3.0	EQUIPMENTANCHORAGE.COM	JONATHAN ROBERSON, S.E.
0.0	Applicant Company Name	Contact Person
	5877 Pine Ave	Suite 210, Chino Hills, CA. 91709
		Mailing Address
	(406) 541-EASE (3273)	jon@easeco.com
	Telephone	E-mail Address
	eby agree to reimburse the Office of States incurred by the department for review.	ewide Health Planning and Development for the actual
		July 5, 2012
	Signa cure of Applicant	
	Principal Engineer	EQUIPMENTANCHORAGE.COM
	Title	Company Name



Office of Statewide Health Planning and Development

4.0	Registered Design Professional Preparing the Report EQUIPMENTANCHORAGE.COM								
=		Company Name							
	Jonathan Roberson, S.E.		S4197						
-	Contact Name								
	5877 Pine Av	e, Suite 210, Chino Hills, CA	. 91709						
=		Mailing Address							
_	909-606-7622		jon@easeco.com						
	Telephone		E-mail Address						
	California Licensed Structural Enginee	r Review and Acceptance of the	he Report						
5.0	EQU								
		Company Name							
_	Jonathan Roberson, S.E.								
	Contact Name		California License Number						
_	5877 Pine Av		. 91709						
		Mailing Address							
_	909-606-7622		jon@easeco.com						
	· · · · · · · · · · · · · · · · · · ·		E-mail Address						
	Anchorage Pre-Approval								
6.0	Anchorage is pre-approved under	OPA-							
	(Separate application for anchorag	e pre-approval is required)							
	Anchorage is not Pre-approved								
-	Certification Method								
70.	☐ Testing in accordance with:	☑ ICC-ES AC-156	Other (Please Specify):						
_	Analysis								
	Experience data								
		and/or Experience Data (Please	Specify):						
-									
_	Testing Laboratory (if applicable)								
3.0	Environmental Testing Laborator	ry. Inc.	Brady Richard						
-	Company Name	<u></u>	Contact Name						
	11034 Indi	EQUIPMENTANCHORAGE.COM Company Name Jonathan Roberson, S.E. Contact Name 5877 Pine Ave, Suite 210, Chino Hills, CA. 91709 Mailing Address 909-606-7622 Telephone Company Name Jonathan Roberson, S.E. Company Name Jonathan Roberson, S.E. Contact Name S877 Pine Ave, Suite 210, Chino Hills, CA. 91709 Mailing Address 909-606-7622 Jonathan Roberson, S.E. S4197 Contact Name S877 Pine Ave, Suite 210, Chino Hills, CA. 91709 Mailing Address 909-606-7622 Telephone Pre-Approval rage is pre-approved under OPA- rate application for anchorage pre-approval is required) Method g in accordance with: ICC-ES AC-156							
-									
	972-247-9657	-	orady@etIdallas.com						
=		_							

 \checkmark

Office of Statewide Health Planning and Development

Approval Param	eters
----------------	-------

9.0	Design in accordance with ASCE 7-05 Chapter 13: Yes No										
	Design Basis of Equipment or Components $(F_p/W_p) = 0.6(S_{DS})(a_P/R_P)(1+2 z/h)$										
	S_{DS} (Spectral response acceleration at short period) = 2.0g @ z/h=1.0 ; 2.6g @ z/h=0										
	a_p (In-structure equipment or component amplification factor) = See Attachment 2										
	R_p (Equipment or component response modification factor) = See Attachment 2										
	I_p (Importance factor) = 1.5										
	z/h (Height factor ratio)= varies										
	Equipment or Component fundamental period(s) = See Attachment 2										
	Building period limits (if any) = NONE										
	Overall dimensions and weight (or range thereof) = See Attachment 1										
	Equipment or Components @ grade designed in accordance with ASCE 7-05 Chapter 15: Yes No										
	Design Basis of Equipment or Components (V/W) =										
	S_{DS} (Spectral response acceleration at short period) =										
	S ₁ (Spectral response acceleration at 1 second period) =										
	R (Response modification coefficient)=1.0										
	Ω_0 (System overstrength factor) =1.0										
	C_d (Deflection amplification factor) =1.0										
	I_p (Importance factor) =1.5										
	Height to Center of Gravity above base =										
	Equipment or Component fundamental period(s) = Sec										
	Overall dimensions and weight (or range thereof) =										
	Tank(s) designed in accordance with ASME BPVC, 2007: Yes No										
10.0	List of attachments supporting the special seismic certification of equipment or components:										
	☐ Test Report ☐ Drawings ☐ Manufacturer's Catalog										
	☐ Calculations ☐ Others (Please Specify): Attachments 1 & 2										
11.0	OSHPD Approval (For Office Use Only)										
_	8/6/2012 December 31, 2016										
	Signature & Date Approval Expiration Date M. R. Karim, SHFR S _{DS} (g) = See Section 9.0										
_	z/h = See Section 9.0										
	Name & Title Special Seismic Certification Valid Up to Condition of Approval (if any):										

TOSHIBA MEDICAL SYSTEMS

SPECIAL SEISMIC CERTIFICATION OF RADIOGRAPHIC IMAGING SYSTEMS

ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS

TABLE 1: SEISMIC CERTIFIED SYSTEMS:

DESCRIPTION	NOTES
Kalare	Special Seismic Certification is limited to the systems identified in Table 1 and the
	subsystem components identified in Table 2.

TABLE 2: SEISMIC CERTIFIED COMPONENTS:

SYSTEM MANUFACTURER	Toshiba Medical Systems								
PRODUCT LINE	Kalare System Components								
SYSTEM			MODEL DIMENSIONS (IN.)				MAX. WT.		
COMPONENT	N	MANUF.	NO.	W	D	Н	(LB.)	MOUNTING	BASIS
Main Processing Ur (MPU)	nit	CMT	HDR-08A	7.8	18.9	24.8	43.0	Rigid Base	UUT2
System Cabinet		TMSC	TA-450F	22.4	15.7	74.8	420	Rigid Base	UUT3
Vertical Wall Stand		QMI	TW-420-T-D	34	40.75 / 42.625	84.19	415	Wall/Floor	UUT1
Vertical Wall Stand		QMI	TW-420-T	34	40.75 / 42.625	84.19	415	Wall/Floor	OSP- 0133-10
Vertical Wall Stand		QMI	TW-420-D	25.19	13.25	84	200	Wall/Floor	INT
Vertical Wall Stand QMI		QMI	TW-420	27.5	12.75	84	225	Wall/Floor	OSP- 0133-10
MOUNTING Rigid Base Mounted (Floor Mounted): a free-standing, base mounted condition with the component rigidly attached to supporting structure and no lateral support above the base Wall/Floor Mounted: a condition where the unit bears on, and is anchored directly to the supporting floor. In addition, lateral restraint anchoring the unit to an adjacent wall or other supporting structure is provided along the height of the equipment. 1. BASIS:						on, lateral			
 1. BASIS: UUT#: Indicates that a test specimen matching these characteristics was tested. SAME: Model is physically, mechanically & electrically the same as test specimen. Difference is limited to model number, color and/or software. NOTES INT (Interpolate): indicates a model that was not specifically tested, and by which seismic qualification was established through evaluation of testing of other, similar models in the product line. MANUF (MANUFACTURER): TMSC = Toshib Medical Systems Corporation CMT = CMT Medical Technologies QMI = Quantum Medical Imaging. 									



TOSHIBA MEDICAL SYSTEMS

SPECIAL SEISMIC CERTIFICATION OF RADIOGRAPHIC IMAGING SYSTEMS

ATTACHMENT 2: TEST SPECIMEN

UUT-1 Wall Stand

MANUFACTURER: Quantum Medical Imaging

MODEL: TW-420-T-D

IDENTIFICATION: S/N: QW420TD-12D-0401

DESCRIPTION: Sub-component of Kalare System.

MOUNTING: Wall/Floor

 a_P 1 R_P 1.5



UUT PROPERTIES:

	DIMENSIONS (in.)		WEIGHT	LOWEST RESONANT FREQUENCY (Hz.)		
WIDTH	DEPTH	HEIGHT	(lb.)	X-Axis	Y-Axis	Z-Axis
34	40.75 / 42.625	83.125	383	6.2	8.0	11.8

UUT-2 Main Processing Unit

MANUFACTURER: CMT Medical Technologies.

MODEL: HDR-08A

IDENTIFICATION: P/N: 84MPU02000

S/N: HRM-0635

DESCRIPTION: Sub-component of Kalare System.

 $\begin{tabular}{lll} MOUNTING: & Floor \\ a_P & 1 \\ R_P & 2.5 \\ \end{tabular}$



UUT PROPERTIES:

	DIN	IENSIONS (in.)		WEIGHT LOWEST RESONANT FRE			JENCY (Hz.)	
WIDTH		DEPTH	HEIGHT	(lb.)	X-Axis	Y-Axis	Z-Axis	
7.8		18.9	24.8	43	10.5	21.9	16.7	





ATTACHMENT PAGE | 2 OF 2

TOSHIBA MEDICAL SYSTEMS

SPECIAL SEISMIC CERTIFICATION OF RADIOGRAPHIC IMAGING SYSTEMS

ATTACHMENT 2: TEST SPECIMEN

UUT-3 System Cabinet

MANUFACTURER: Toshiba

MODEL: TA-450F

IDENTIFICATION: SN: 1234567890

DESCRIPTION: Sub-component of Kalare System.

MOUNTING: Wall/Floor

 a_P 2.5 R_P 6.0



UUT PROPERTIES:

DIN	MENSIONS (in.)		WEIGHT	WEIGHT LOWEST RESONANT FREQUENCY (H	HT LOWEST RESONANT FREQUENCY (Hz.)	NCY (Hz.)
WIDTH	DEPTH	HEIGHT	(lb.)	X-Axis	Y-Axis	Z-Axis
22.4	15.7	74.8	418	12.6	20.4	34.5