



**OFFICE OF STATEWIDE HEALTH PLANNING AND DEVELOPMENT  
FACILITIES DEVELOPMENT DIVISION**

**APPLICATION FOR OSHPD SPECIAL SEISMIC  
CERTIFICATION PREAPPROVAL (OSP)**

OFFICE USE ONLY

**APPLICATION #: OSP-0421**

**OSHPD Special Seismic Certification Preapproval (OSP)**

Type:  New  Renewal

**Manufacturer Information**

Manufacturer: Siemens Healthcare GmbH Manufacturer's

Technical Representative: Stephan Maurer

Mailing Address: Siemensstr. 3, 91301 Forchheim, Germany

Telephone: +49 (9191) 18-8835

Email: Stephan.sm.maurer@siemens-healthineers.com

**Product Information**

Product Name: CT Systems

Product Type: NA

Product Model Number: See Attachment 1, Table 1

General Description: Multiple component system for the provision of Computed Tomography medical diagnostic applications

Mounting Description: Rigid floor mounted and combined rigid floor/wall mounted

Tested Seismic Enhancements: None

DATE: 07/12/2021

**Applicant Information**

Applicant Company Name: W.E. Gundy & Associates, Inc.

Contact Person: Travis Soppe

Mailing Address: 1199 Shoreline Drive, Boise, ID 83702

Telephone: (208) 342-5989

Email: tsoppe@wegai.com

Title: President





**OFFICE OF STATEWIDE HEALTH PLANNING AND DEVELOPMENT  
FACILITIES DEVELOPMENT DIVISION**

**California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)**


Company Name: W.E. GUNDY & ASSOCIATES INC.  
Name: Travis Soppe California License Number: S6115  
Mailing Address: P.O. Box 9121, Boise, ID 83707  
Telephone: (208) 342-5989 Email: tsoppe@wegai.com

**Certification Method**

GR-63-Core       ICC-ES AC156       IEEE 344       IEEE 693       NEBS 3  
 Other (Please Specify): \_\_\_\_\_

**Testing Laboratory**

Company Name: ENVIRONMENTAL TESTING LABORATORIES, INC. (ETL)  
Contact Person: Brady Richard  
Mailing Address: 11034 Indian Trail, Dallas TX 75229-3513  
Telephone: (972) 247-9657 Email: brady@etldallas.com





**OFFICE OF STATEWIDE HEALTH PLANNING AND DEVELOPMENT  
FACILITIES DEVELOPMENT DIVISION**

**Seismic Parameters**

Design Basis of Equipment or Components ( $F_p/W_p$ ) = See attachments

SDS (Design spectral response acceleration at short period, g) = 2.0

$a_p$  (Amplification factor) = See attachments

$R_p$  (Response modification factor) = See attachments

$\Omega_0$  (System overstrength factor) = See Attachment

$I_p$  (Importance factor) = 1.5

$z/h$  (Height ratio factor) = 1

Natural frequencies (Hz) = See attachments

Overall dimensions and weight = See attachments


**OSHPD Approval (For Office Use Only) - Approval Expires on 12/31/2025**

Date: 7/12/2021

Name: Mohammad Aliaari Title: Senior Structural Engineer

Special Seismic Certification Valid Up to: SDS (g) = 2.0 z/h = 1

Condition of Approval (if applicable): DATE: 07/12/2021



<b>TABLE 1</b>	<b>SIEMENS HEALTHCARE GmbH</b>					 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC.</small> <small>STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>		
	<b>SPECIAL SEISMIC CERTIFICATION</b>							
<b>CERTIFIED SYSTEM AND COMPONENTS</b>								

**Manufacturer:** Siemens Healthcare GmbH

**System:** SOMATOM Force CT

System Component <sup>1)</sup>	Siemens Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT
		Width	Depth	Height			

**Gantries**

SOMATOM Force	10742327	94.5	46.5	78.4	5880	floor	UUT-1
---------------	----------	------	------	------	------	-------	-------

**Patient Tables**

PHS-5	10742323	29.5	100.5 - 179.5	24.3 - 41.7	1066	floor	UUT-2 <sup>2</sup>
MPT4	10742324	27.8	99.5 - 179.5	23.0 - 38.5	1388	floor	UUT-3 <sup>2</sup>

**Power Distribution Cabinet**

Power Distribution Cabinet	10757110	47.2	29.5	75.2	2027	floor/wall	UUT-4
----------------------------	----------	------	------	------	------	------------	-------

**Image Control and Reconstruction**

Image Control System	10864203	17.3	18.9	16.9	34	floor	UUT-5
Image Reconstruction System	10742951	12.2	27.0	19.7	86	floor	UUT-6

<sup>1</sup> All components are manufactured by Siemens Healthcare GmbH unless noted. Part numbers listed uniquely identify type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> Patient table weight does not include simulated patient weights of 440lb.

**SEISMIC CERTIFICATION LIMITS**

System Component	Code	S <sub>Ds</sub> (g)	z / h	I <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>	F <sub>P</sub> / W <sub>P</sub>
Gantries	CBC 2019	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Patient Tables	CBC 2019	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Power Distribution Cabinet	CBC 2019	2.0	1.0	1.50	2.5	6.0	2.0	1.50
Image Control and Reconstruction	CBC 2019	2.0	1.0	1.50	1.0	2.5	2.0	1.44

**UUT-1**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" grade 8 bolts



**Manufacturer:** Siemens Healthcare GmbH      **Test Location:** Environmental Testing Laboratory

**Component:** SOMATOM Force Gantry      **Test Date:** November 2014

**Model Number:** 10742327      **Report Number:** SQ35-1423-01

**UUT Function:** Continuous rotating detector for high-resolution data acquisition

**UUT Description:** Component of the SOMATOM Force CT system - water cooled

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
5,880	94.5	46.5	78.4	6.7	11.2	10.2

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

**UUT-2**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" grade 8 bolts.



BY: Mohammad Aliaari

DATE: 07/12/2021

<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> PHS-5 Patient Table	<b>Test Date:</b> November 2014
<b>Model Number:</b> 10742323	<b>Report Number:</b> SQ35-1423-01
<b>UUT Function:</b> Motorized patient table	
<b>UUT Description:</b> Component of SOMATOM CT Systems	

**UUT PROPERTIES**

Weight (lb) with Patient*	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,506	29.5	100.5 - 179.5	24.3 - 41.7	2.9	2.6	10.7

\*The patient table moves vertically and horizontally to accommodate different patients and procedures. The system was tested in the normal vertical operating position, with the table top extended, and a total simulated patient weight of 440lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

**UUT-3**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with 4 - 5/8" grade 8 bolts.



<b>Manufacturer:</b> Siemens Healthcare GmbH	<b>Test Location:</b> Environmental Testing Laboratory
<b>Component:</b> MPT4 Patient Table	<b>Test Date:</b> November 2014
<b>Model Number:</b> 10742324	<b>Report Number:</b> SQ35-1423-01
<b>UUT Function:</b> Motorized patient table	
<b>UUT Description:</b> Component of SOMATOM CT Systems	

**UUT PROPERTIES**

Weight (lb) with Patient*	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,828	27.8	99.5 - 179.5	23.0 - 38.5	2.7	3.3	7.7

\*The patient table moves vertically and horizontally to accommodate different patients and procedures. The system was tested in the normal vertical operating position, with the table top extended, and a total simulated patient weight of 440lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

**UUT-4**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid wall mounted at top back side of unit with 4 - 5/16" bolts and rigid floor mounted with 2 - 1/2" grade 8 bolts. Wall mount brackets constructed of 4" long L4x2.5x0.25 angles that bolt thru the short leg to the UUT with a single 1/2" grade 8 bolt.



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Power Distribution Unit | **Test Date:** November 2014

**Model Number:** 10757110 | **Report Number:** SQ35-1423-01

**UUT Function:** Power distribution to CT system

**UUT Description:** Component of SOMATOM CT Systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
2,027	47.2	29.5	75.2	13.2	12.7	>33

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.



**UUT-5**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with two hand-tightened, 1" wide straps (200lb WLL), thru L2.5x2.5x1/4" x 2.5" brackets positioned on either side of the unit (spaced 12" apart). The four angle brackets are attached to the table with (1) 3/8" bolt at each bracket location.



**Manufacturer:** Siemens Healthcare GmbH      **Test Location:** Environmental Testing Laboratory

**Component:** Image Control System      **Test Date:** November 2014

**Model Number:** 10864203      **Report Number:** SQ35-1423-01

**UUT Function:** Computer for data acquisition, image reconstruction, and processing

**UUT Description:** Component of SOMATOM CT Systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
34	7.3	18.9	16.9	>33	30.2	>33

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.

**UUT-6**

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with two hand-tightened, 1" wide straps (200lb WLL), thru L2.5x2.5x1/4" x 2.5" brackets positioned on either side of the unit (spaced 12" apart). The four angle brackets are attached to the table with (1) 3/8" bolt at each bracket location.



**Manufacturer:** Siemens Healthcare GmbH | **Test Location:** Environmental Testing Laboratory

**Component:** Image Reconstruction System (IRSmx4) | **Test Date:** November 2014

**Model Number:** 10742951 | **Report Number:** SQ35-1423-01

**UUT Function:** Computer for data acquisition, image reconstruction, and processing

**UUT Description:** Component of SOMATOM CT Systems

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
86	12.2	27.0	19.7	17.5	14.6	>33

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2019 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	1.34	0.54

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 Test.