



DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION
FACILITIES DEVELOPMENT DIVISION

APPLICATION FOR HCAI SPECIAL SEISMIC
CERTIFICATION PREAPPROVAL (OSP)

OFFICE USE ONLY

APPLICATION #: OSP-0628

HCAI Special Seismic Certification Preapproval (OSP)

Type: ☐ New ☒ Renewal

Manufacturer Information

Manufacturer: GE Healthcare

Manufacturer's Technical Representative: Tom Farrow

Mailing Address: 3000 N. Grandview Blvd., Waukesha, WI 531881696

Telephone: (888) 406-1101

Email: Tom.Farnow@gehcseismic.com

Product Information

Product Name: CT Systems

Product Type: NA

Product Model Number: See Attachment 1

General Description: System components of multiple-component medical diagnostic imaging systems. Special Seismic Certification is limited to components identified in Attachment 1 for functions related to Computed Tomography (CT) imaging services only.

Mounting Description: Rigid Base mount

Tested Seismic Enhancements: Seismic enhancements made to the test units and/or modifications required to address anomalies during the tests shall be incorporated into the production units.

Applicant Information

Applicant Company Name: EASE

Contact Person: Jonathan Roberson

Mailing Address: 5877 Pine Ave, Suite 210, Chino Hills, CA 91709

Telephone: (909) 606-7622

Email: j.roberson@easeco.com

Title: SE



**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION
FACILITIES DEVELOPMENT DIVISION**

California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)

Company Name: EASE LLC

Name: Jonathan Roberson

California License Number: S4197

Mailing Address: 5877 Pine Ave., Suite 210, Chino Hills, CA 91709

Telephone: (951) 295-1892

Email: jon@EASECo.com

Certification Method

☐ GR-63-Core

☒ ICC-ES AC156

☐ IEEE 344

☐ IEEE 693

☐ NEBS 3

☐ Other (Please Specify): _____

Testing Laboratory


Company Name: ENVIRONMENTAL TESTING LABORATORIES, INC. (ETL)

Contact Person: Jeremy Lange

Mailing Address: 11034 Indian Trail, Dallas TX 75229-3513

Telephone: (972) 247-9657

Email: Jeremy@etldallas.com





DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION FACILITIES DEVELOPMENT DIVISION

Seismic Parameters

Design Basis of Equipment or Components (F_p/W_p) = See Attachment 1

SDS (Design spectral response acceleration at short period, g) = 2.00 (z/h = 1); 2.50 (z/h = 0)

a_p (Amplification factor) = See Attachment 1

R_p (Response modification factor) = See Attachment 1

Ω_0 (System overstrength factor) = See Attachment

I_p (Importance factor) = 1.5

z/h (Height ratio factor) = 1 and 0

Natural frequencies (Hz) = See Attachment 2

Overall dimensions and weight = See Attachment 1

HCAI Approval (For Office Use Only) - Approval Expires on 07/10/2029

Date: 7/10/2023

Name: Mohammad Karim

Title: Supervisor, Health Facilities

Special Seismic Certification Valid Up to: SDS (g) = See Above

z/h = See Above

Condition of Approval (if applicable):

ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS

ATTACHMENT PAGE | 1 OF 1

TABLE 1:

SYSTEM MFR.	GE Healthcare													
SYSTEM	Revolution Maxima / Revolution Ace													
COMPONENT	MODEL NUMBER	APPROX. DIMENSIONS (IN.)			MAX. WT. (LB.)	MOUNT ^[6]	BASIS ^[1]	F _P /W _P	S _{DS}	z/h	a _P	R _P	Ω ₀	
		W	D	H										
GANTRIES ^[5]														
Revolution Maxima Gantry	6670000-2	80.4	40.2	76.3	3434	Floor	UUT1914-1	2.40 1.13	2.0 2.5	1 0	1	1 ½	1 ½	
Revolution Maxima Gantry	6670000-4	80.4	40.2	76.3	3434	Floor	SAME							
Revolution Maxima Gantry	6670000-200	80.4	40.2	76.3	3434	Floor	SAME							
Revolution Maxima Gantry	6670000-400	80.4	40.2	76.3	3434	Floor	SAME							
Revolution Ace Gantry	6670000-3	80.4	40.2	76.3	3434	Floor	SAME							
Revolution Ace Gantry	6670000-5	80.4	40.2	76.3	3434	Floor	SAME							
PATIENT TABLES ^[5]														
GT1700V Table	5122080-11 ^[4]	25.6	93.3	19.2 / 41.2	1059 ^[2]	Floor	UUT1126-2	2.40 1.13	2.0 2.5	1 0	1	1 ½	1 ½	
GT1700V Table	5122080-12	25.6	93.3	19.2 / 41.2	1059	Floor	SAME							
CONSOLES														
Open Console w/ Z8G4	5946404-15	15.8	26.4	22.7	142 ^[3]	Floor	UUT1914-2	1.44 1.13	2.0 2.5	1 0	1	2 ½	2	
Open Console w/ Z8G4	5946404-25	15.8	26.4	22.7	142 ^[3]	Floor	SAME							
POWER DISTRIBUTION														
Power Distribution Unit	2326492-80 ^[4]	27.6	21.7	41.8	841 ^[3]	Floor	UUT1126-3	1.44 1.13	2.0 2.5	1 0	1	2 ½	2	
Power Distribution Unit	2326492-81	27.6	21.7	41.8	841 ^[3]	Floor	SAME							
MOUNT	Floor (Rigid Base): free-standing, base-mounted tower configuration with the component rigidly attached to a supporting structure and no lateral support above the base.													
NOTES	<div><div>1. BASIS</div><div><div>• UUT#: Indicates that a test specimen matching these characteristics was tested as part of this testing program.</div><div>• SAME: Model is physically, mechanically & electrically identical to a tested unit, except for possible variations in model number, color and/or software.</div></div><div>2. 1700V Patient Table weight does not include the 350 lb. simulated patient load present during testing.</div><div>3. Weight includes mounting bracket hardware.</div><div>4. Model is not used with the system(s) identified in this table; it is included only to establish basis for seismic qualification.</div><div>5. Requires use of additional GE bushing kit B76632DA (P/N 5841169).</div><div>6. See Attachment 2 for additional mounting information.</div></div>													

ATTACHMENT 2: TEST SPECIMEN SUMMARY

ATTACHMENT PAGE | 1 OF 2

UUT 1914-1 Revolution Maxima CT Gantry						
Manufacturer:		GE Hangwei Medical Systems Co., LTD.				
Identification:		Model No.: 6670000-2 Serial No.: 000000HM11				
Description:		Component of Revolution Maxima - Non-tilt Gantry - Touchscreen display on front (x2) - Configuration controlled verification unit that is equivalent to standard production.				
Mounting:		Rigid Base (Floor) mounted using: (4) – 5/8" dia. SAE J429 Grade 8 bolts to test fixture. GE Bushing Kit P/N 5841169.				
Dimensions (in.)				Lowest Resonant Frequency (Hz.)		
Width	Depth	Height	Weight (lb.)	Front-Axis	Side-Axis	Vertical-Axis
80.4	40.2	76.3	3434	9.9	11.6	14.0
ICC-ES AC156 Shake Table Test Parameters						Code: 2022 CBC
S _{DS} (G)	z/h	I _p	A _{FLX-H} (G)	A _{RIG-H} (G)	A _{FLX-V} (G)	A _{RIG-V} (G)
2.0 2.5	1 0	1.5	3.2	2.4	1.68	0.68
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.						



UUT 1914-2 Open Console w/ Z8G4						
Manufacturer:		GE Hangwei Medical Systems Co., LTD.				
Identification:		Model No.: 5946404-15 Serial No.: 000000HM1				
Description:		Component of the Revolution Maxima / Revolution Ace CT Systems - HP Z8G4 Computer - Open style console - Configuration controlled verification unit that is equivalent to standard production.				
Mounting:		Rigid Base (Floor) mounted using: GE Console Anchor Kit 5812703-2 (3) – 3/8" dia. SAE J429 Grade 8 bolts to test fixture. NOTE: Console Anchor Kit 5812703-2 is included standard w/ console.				
Dimensions (in.)				Lowest Resonant Frequency (Hz.)		
Width	Depth	Height	Weight (lb.)	Side-Axis	Front-Axis	Vertical-Axis
15.8	26.4	22.7	141.5 w/ brackets	14.6	30.1	26.2
ICC-ES AC156 Shake Table Test Parameters						Code: 2022 CBC
S _{DS} (G)	z/h	I _p	A _{FLX-H} (G)	A _{RIG-H} (G)	A _{FLX-V} (G)	A _{RIG-V} (G)
2.0 2.5	1 0	1.5	3.2	2.4	1.68	0.68
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.						



ATTACHMENT 2: TEST SPECIMEN SUMMARY

ATTACHMENT PAGE | 2 OF 2

UUT 1126-2 GT1700V Table						
Manufacturer:		GE Hangwei Medical Systems Co., LTD				
Identification:		Model No.: 5122080-11				
Description:		<p>System component of the Optima CT660 System. Also used with other CT systems. Test specimen included a simulated patient load of 350 lb.</p> <p>Requires additional kit as follows:</p> <ul style="list-style-type: none">• B7660MY (Optima CT660 system)• B76632DA (Revolution Maxima / Revolution Ace)				
Mounting:		<u>Rigid Base (Floor) mounted using</u> (4) – 5/8" dia. hex head bolts to test fixture.				
Dimensions (in.)				Lowest Resonant Frequency (Hz.)		
Width	Depth	Height	Weight (lb.)	Transverse-Axis	Longitudinal-Axis	Vertical-Axis
25.6	93.3	19.2 / 41.2	1059+ 350 Patient	3.9	15.2	14.2
ICC-ES AC156 Shake Table Test Parameters						Code: 2022 CBC
S _{DS} (G)	z/h	I _P	A _{FLX-H} (G)	A _{RIG-H} (G)	A _{FLX-V} (G)	A _{RIG-V} (G)
2.0	1	1.5	3.2	2.4	1.74	0.70
2.6	0					
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.						



UUT 1126-3 Power Distribution Unit (PDU) 07/10/2023						
Manufacturer:		GE Hangwei Medical Systems Co., LTD.				
Identification:		Model No.: 2326492-80 Serial No.: 270337HM9				
Description:		<p>System component of the Optima CT660 System. Also used with other CT systems.</p>				
Mounting:		<p><u>Rigid Base (Floor) mounted using:</u> (4) – 3/8" dia. ASTM A574 Socket Head Cap Screws w/ washer through GE mounting brackets to floor plate. (2 anchors ea. bracket)</p> <p><u>GE mounting assembly including:</u> (2) – Seismic Bracket, System Cabinet (P/N 2354563-2) Each bracket mounted to cabinet w/: (2) – M10 x 25mm Class 12.9 bolt (Torque= 38.4 N-m) (2) – M10 Lock Washer (P/N 2203-M10-07) (2) – M10 Flat Washer (P/N 2000-M10-03)</p>				
Dimensions (in.)				Lowest Resonant Frequency (Hz.)		
Width	Depth	Height	Weight (lb.)	Front-Axis	Side-Axis	Vertical-Axis
27.6	21.7	41.8	841	20.4	20.6	13.8
ICC-ES AC156 Shake Table Test Parameters						Code: 2022 CBC
S _{DS} (G)	z/h	I _P	A _{FLX-H} (G)	A _{RIG-H} (G)	A _{FLX-V} (G)	A _{RIG-V} (G)
2.0	1	1.5	3.2	2.4	1.74	0.70
2.6	0					
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.						

