



**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT**

**APPLICATION FOR HCAI SPECIAL SEISMIC  
CERTIFICATION PREAPPROVAL (OSP)**

OFFICE USE ONLY

**APPLICATION #: OSP-0631**

**HCAI Special Seismic Certification Preapproval (OSP)**

Type:  New  Renewal

**Manufacturer Information**

Manufacturer: Canon Medical Systems Corporation

Manufacturer's Technical Representative: Krystal Ouren

Mailing Address: 2441 Michelle Drive, Tustin, CA 92780

Telephone: (949) 540-5781

Email: kouren@us.medical.canon

**Product Information**

Product Name: See Attachment 1, Table 1

Product Model Number(s): See Attachment 1, Table 1

Product Category: CT Systems

Product Sub-Category: CT Systems

General Description: Multiple component systems for the provision of Computed Tomography (CT) medical imaging. Seismic Certification is limited to the systems & components identified in Attachment 1 for diagnostic assessment of trauma.

Mounting Description: Rigid base mounted (i.e. floor mounted), except as noted in Attachment 1, Table 1.

Tested Seismic Enhancements: Seismic enhancements made to the test units and/or modifications required to address anomalies during the tests shall be incorporated into the production units.

**Applicant Information**

Applicant Company Name: EASE

Contact Person: Jonathan Roberson

Mailing Address: 5877 Pine Ave, Suite 210, Chino Hills, CA 91709

Telephone: (909) 606-7622

Email: j.roberson@easeco.com

Title: SE





**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT**

**California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)**

Company Name: EASE LLC

Name: Jonathan Roberson California License Number: S4197

Mailing Address: 5877 Pine Ave., Suite 210, Chino Hills, CA 91709

Telephone: (951) 295-1892 Email: jon@EASECo.com

**Certification Method**

- GR-63-Core
- ICC-ES AC156
- IEEE 344
- IEEE 693
- NEBS 3
- Other (Please Specify): \_\_\_\_\_


**Testing Laboratory**

Company Name: ENVIRONMENTAL TESTING LABORATORIES, INC. (ETL)

Contact Person: Jeremy Lange

Mailing Address: 11034 Indian Trail, Dallas TX 75229-3513

Telephone: (972) 247-9657 Email: jeremy@etldallas.com






**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT**

**Seismic Parameters**

Design Basis of Equipment or Components ( $F_p/W_p$ ) =	See Attachment 1, Table 1
SDS (Design spectral response acceleration at short period, g) =	See Attachment 1, Table 1
$a_p$ (Amplification factor) =	See Attachment 1, Table 1
$R_p$ (Response modification factor) =	See Attachment 1, Table 1
$\Omega_0$ (System overstrength factor) =	See Attachment
$I_p$ (Importance factor) =	1.5
$z/h$ (Height ratio factor) =	0 and 1
Natural frequencies (Hz) =	See Attachment 2
Overall dimensions and weight =	See Attachment 2

<b>HCAI Approval (For Office Use Only) - Approval Expires on 03/09/2032</b>	
Date: <u>3/9/2026</u>	
Name: <u>Mohammad Karim</u>	Title: <u>Supervisor, Health Facilities</u>
Special Seismic Certification Valid Up to: SDS (g) = <u>See Attachment 1, Table 1</u>	$z/h$ = <u>1</u>
Condition of Approval (if applicable): _____	



**ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS**

ATTACHMENT PAGE | 1 OF 3

**TABLE 1:**

Manufacturer		CANON MEDICAL SYSTEMS CORPORATION							S <sub>Ds</sub> = 2.0 at z/h = 1		
System		Aquilion Prime SP (TSX-303B/8)							S <sub>Ds</sub> = 2.5 at z/h = 0		
Component	Model No.	Approx. Dimensions [in]			Wt. [lb]	Mount	Basis <sup>[1]</sup>	F <sub>P</sub> /W <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>
		W	D	H							
<b>CT GANTRIES</b>											
Aquilion CT Scanner Gantry - Type D	CGGT-032A/3A	84.7	34.25	73	4061	Rigid Base	UUT-1703-4	2.40 1.13	1	1 ½	2
w/ Optional Gantry Rear Operating Panel Kit	CAGP-002A/3B										
Aquilion CT Scanner Gantry - Type D	CGGT-032A/4A	89.4	37.8	75.8	5171	Rigid Base	INT	2.40 1.13	1	1 ½	2
w/ Optional Gantry Rear Operating Panel Kit	CAGP-002A/3B										
Aquilion CT Scanner Gantry - Type G	CGGT-039A/1A	89.4	37.8	75.8	5171	Rigid Base	UUT-1906-2	2.40 1.13	1	1 ½	2
w/ Optional Gantry Rear Operating Panel Kit	CAGP-003A/2B										
w/ Optional Area Finder	CGAP-001A/1B										
<b>PATIENT COUCHES</b>											
Patient Couch - High Capacity – Extended <sup>[2]</sup>	CBTB-032A/1A	26	113.8 - 197.7	18.5 - 41.0	1512 <sup>[4]</sup>	Rigid Base	UUT-1303-2A	2.40 1.13	1	1 ½	2
Patient Couch - High Capacity – Extended <sup>[2]</sup>	CBTB-032A/8A	26	113.8 - 197.7	18.5 - 41.0	1512	Rigid Base	SAME AS UUT-1303-2A	2.40 1.13	1	1 ½	2
Patient Couch – High Capacity – Extended	CBTB-032A/8A	26.0	113.8 - 197.7	18.5 - 41.0	1950 <sup>[4]</sup>	Rigid Base	UUT-2006-1	2.40 1.13	1	1 ½	2
w/ Optional Lateral Movement Kit	CALU-001A/6B										
w/ Optional CT Fluoro <sup>[3]</sup>	TSXF-004A/3B										
Patient Couch - High Capacity - Extended	CBTB-032A/8A	26	113.8 - 197.7	18.5 - 41.0	1512	Rigid Base	INT	2.40 1.13	1	1 ½	2
w/ Optional Lateral Movement Kit	CALU-001A/6B	---	---	3.5	396						
Patient Couch - High Capacity - Compact	CBTB-032B/8A	26	94.1 - 158.4	18.5 - 41.0	1445	Rigid Base	INT	2.40 1.13	1	1 ½	2
w/ Optional Lateral Movement Kit	CALU-001A/6B	---	---	3.5	396						
Patient Couch - High Capacity – Compact <sup>[2]</sup>	CBTB-032B/8A	26	94.1 - 158.4	18.5 - 41.0	1445	Rigid Base	INT	2.40 1.13	1	1 ½	2
Patient Couch – High Capacity – Compact	CBTB-032B/8A	26.0	94.1 - 158.4	22.3 - 41	1893	Rigid Base	INT	2.40 1.13	1	1 ½	2
w/ Optional Lateral Movement Kit	CALU-001A/6B										
w/ Optional CT Fluoro <sup>[3]</sup>	TSXF-004A/4B										

Table continued next page

**ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS**

ATTACHMENT PAGE | 2 OF 3

**TABLE 1:**

Manufacturer		CANON MEDICAL SYSTEMS CORPORATION										S <sub>DS</sub> = 2.0 at z/h = 1	
System		Aquilion Prime SP (TSX-303B/8)										S <sub>DS</sub> = 2.5 at z/h = 0	
Component	Model No.	Approx. Dimensions [in]			Wt. [lb]	Mount	Basis <sup>[1]</sup>	F <sub>P</sub> /W <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>		
		W	D	H									
<b>PATIENT COUCHES (continued)</b>													
Patient Couch – High Capacity – Compact <sup>[2]</sup>	CBTB-032B/8A	26.0	94.1 - 158.4	18.5 - 41	1497	Rigid Base	INT	2.40 1.13	1	1 ½	2		
w/ Optional CT Fluoro <sup>[3]</sup>	TSXF-004A/4B												
Patient Couch – High Capacity – Extended <sup>[2]</sup>	CBTB-032A/8A	26.0	113.8 - 197.7	18.5 - 41.0	1554 <sup>[4]</sup>	Rigid Base	UUT-2006-2	2.40 1.13	1	1 ½	2		
w/ Optional CT Fluoro <sup>[3]</sup>	TSXF-004A/3B												
<b>CONSOLE COMPONENTS</b>													
Con Box - Type B <sup>[5]</sup>	CKCN-020C/1B	23.2	37.6	53.7	617	Rigid Base	UUT-2002-7	1.44 1.13	1	2 ½	2		
Con Box - Type B <sup>[5]</sup>	CKCN-020C/1B	23.2	37.6	53.7	651	Rigid Base	UUT-2002-8	1.44 1.13	1	2 ½	2		
w/ Optional Display Console Kit	CGS-72B/1B												
NAVI Box – Type A <sup>[6]</sup>	CKCN-017B/2B	7.9	12.2	13.8	28	CTA	UUT-1703-5	1.44 1.13	1	2 ½	2		
NAVI Box – Type A <sup>[6]</sup>	CKCN-020C/1B	7.9	12.2	13.8	28	CTA	SAME AS UUT-1703-5	1.44 1.13	1	2 ½	2		
<b>POWER DISTRIBUTION</b>													
Power Distributor - Type D	CETF-006C/2A	33.5	26.8	48.8	1238	Rigid Base	UUT-1703-7	1.44 1.13	1	2 ½	2		
w/ Optional UPS Kit for Power Distributor	CEUC-001B/4B												
<b>USER INTERFACE</b>													
Monitor (Mfr: EIZO)	MX-192	16.3	8.1	16.1 - 21.5	13	CTA	UUT-1603-6	1.44 1.13	1	2 ½	2		
Keyboard (scan)	BSX74-2866-01*B	21	9.25	3.0	7	CTA	UUT-1901-8	1.44 1.13	1	2 ½	2		
Keyboard (scan)	BSX74-2866-01*D	21	9.25	3.0	7	CTA	SAME AS UUT-1901-8	1.44 1.13	1	2 ½	2		
Keyboard (display)	BSX74-2866-02*A	21	9.25	3.0	7	CTA	UUT-1603-13	1.44 1.13	1	2 ½	2		
Keyboard (display)	BSX74-2866-02*D	21	9.25	3.0	7	CTA	SAME AS UUT-1603-13	1.44 1.13	1	2 ½	2		

Table continued next page

**ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS**

**TABLE 1:**

<i>Manufacturer</i>	CANON MEDICAL SYSTEMS CORPORATION										$S_{DS} = 2.0$ at $z/h = 1$	
<i>System</i>	Aquilion Prime SP (TSX-303B/8)										$S_{DS} = 2.5$ at $z/h = 0$	
Component	Model No.	Approx. Dimensions [in]			Wt. [lb]	Mount	Basis <sup>[1]</sup>	$F_p/W_p$	$a_p$	$R_p$	$\Omega_0$	
		W	D	H								
<i>Mount</i>	<u>Rigid Base</u> : free-standing, base-mounted configuration with the component rigidly attached to a supporting structure and no lateral support above the base. <u>CTA (Countertop Anchored)</u> : refers to a condition where the unit is anchored to a counter, desk, or other piece of rigid fixed furniture.											
<i>Notes</i>	<ol style="list-style-type: none"> <li>BASIS:             <ul style="list-style-type: none"> <li>"UUT-#" indicates that a test specimen matching these characteristics was tested as part of a seismic test program.</li> <li>"SAME" indicates component is physically, mechanically, and electrically the same as another test specimen with differences limited to model number, color, and/or software.</li> <li>"INT" (Interpolated or Extrapolated) indicates model of equipment not specifically tested for which seismic qualification is established through testing of similar models in the product line.</li> </ul> </li> <li>Patient Couch – High Capacity (without Lateral Kit) requires the use of Seismic Kit P/N: ANC-CA-CBTB032</li> <li>CT Fluoro Option requires the use of seismic kit Model No. PX79-68940-1</li> <li>Patient Couch weight does not include 660 lb Max Patient Load</li> <li>Con Box – Type B requires the use of Seismic Kit P/N: ANC-CA-002COX</li> <li>NAVI Box – Type A requires the use of seismic kit P/N: ANC-CA-002NAVI</li> <li><i>Gray italic formatting indicates unit is not a component of system and used as basis for seismic qualification only.</i></li> </ol>											

