



**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**APPLICATION FOR HCAI SPECIAL SEISMIC  
CERTIFICATION PREAPPROVAL (OSP)**

OFFICE USE ONLY

**APPLICATION #: OSP-0765**

**HCAI Special Seismic Certification Preapproval (OSP)**

Type:  New  Renewal

**Manufacturer Information**

Manufacturer: Shanghai United Imaging Healthcare Co., Ltd.

Manufacturer's Technical Representative: Xin GAO

Mailing Address: 2258 Chengbei Rd., Jiading District, Shanghai, 201807

Telephone: +86 (21) 67076888

Email: xin.gao@united-imaging.com

**Product Information**

Product Name: CT Systems

Product Type: NA

Product Model Number: ATLAS CT System

General Description: Multiple component system used for producing Computed Tomography (CT) medical images for diagnostic results.

Mounting Description: Rigid, Floor Mounted

Tested Seismic Enhancements: Seismic enhancements made to the test units and/or modifications required to address anomalies during the tests shall be incorporated into the production units.

**Applicant Information**

Applicant Company Name: WE Gundy & Associates, Inc


Contact Person: Travis Soppe

Mailing Address: PO Box 9121, Boise, ID 83707

Telephone: (208) 342-5989

Email: tsoppe@wegai.com

Title: President





**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)**


Company Name: W.E. GUNDY & ASSOCIATES INC.  
Name: Travis Soppe California License Number: S6115  
Mailing Address: P.O. Box 9121, Boise, ID 83707  
Telephone: (208) 342-5989 Email: tsoppe@wegai.com

**Certification Method**

GR-63-Core       ICC-ES AC156       IEEE 344       IEEE 693       NEBS 3  
 Other (Please Specify): \_\_\_\_\_

**Testing Laboratory**

Company Name: SUZHOU ELECTRICAL APPARATUS SCIENCE RESEARCH INSTITUTE CO., LTD. (EETI)  
Contact Person: Tangfu Ji  
Mailing Address: No. 5, Yuexi Qianzhu Road, Wuzhong District, Suzhou, 215104 Jiangsu, CHINA  
Telephone: 86 (512) 69552195 Email: eeti-kz@eeti.cn







**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
FACILITIES DEVELOPMENT DIVISION**

**Seismic Parameters**

Design Basis of Equipment or Components ( $F_p/W_p$ ) =	See Attachments
SDS (Design spectral response acceleration at short period, g) =	2.00 (z/h = 1.0) and 2.50 (z/h = 0.0)
$a_p$ (Amplification factor) =	See attachments
$R_p$ (Response modification factor) =	See attachments
$\Omega_0$ (System overstrength factor) =	See Attachment
$I_p$ (Importance factor) =	1.5
z/h (Height ratio factor) =	1 and 0
Natural frequencies (Hz) =	See Attachment
Overall dimensions and weight =	See Attachment

<b>HCAI Approval (For Office Use Only) - Approval Expires on 09/20/2029</b>			
Date:	9/20/2023	OSP-0765	
Name:	Mohammad Karim	BY: Mohammad Karim	Title: Supervisor, Health Facilities
Special Seismic Certification Valid Up to: SDS (g) =	See Above	z/h =	See Above
Condition of Approval (if applicable):	DATE: 09/20/2023		



<b>Table 1</b>	<b>UNITED IMAGING HEALTHCARE Co., Ltd SPECIAL SEISMIC CERTIFICATION CERTIFIED SYSTEM AND COMPONENTS</b>	 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC. STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>
----------------	---	---

<b>System:</b> ATLAS uCT Systems	<b>Manufacturer:</b> United Imaging Healthcare Co., Ltd
----------------------------------	---

System Component <sup>1</sup>	United Imaging Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT <sup>3</sup>
		Width	Length	Height			
<b>Gantries</b>							
uCT ATLAS	7808924	91.8	36.6	77.8	5732	floor	UUT <sub>x</sub> -9
uCT ATLAS Ascend	78009761	91.8	36.6	77.8	5688	floor	interpolated
uCT ATLAS Astound	7809764	91.8	36.6	77.8	5600	floor	interpolated
uCT 760/780	78001101	91.7	31.5	77.8	5157	floor	UUT <sub>y</sub> -2
<b>Patient Table<sup>2</sup></b>							
uCT - PS2	80100042	23.6	119.9-205.8	18.9-37.4	1060	floor	UUT <sub>y</sub> -5
uCT - PS3	78007207	23.6	120.0-205.8	18.9-37.4	1060	floor	interpolated
uCT - PS4	78008392	22.1	119.5-183.8	18.9-37.4	2095	floor	UUT <sub>x</sub> -10
<b>Power Supply Cabinet</b>							
PSC - ATLAS	78004466	27.6	29.6	59.5	1390	floor	UUT <sub>x</sub> -11

**Notes:**

<sup>1</sup> All components are manufactured by United Imaging Healthcare Co., Ltd. The part numbers listed uniquely identify the type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> Patient table weights do not include 362lb simulated patient weight for UUT<sub>y</sub>-5 and 565lbs for UUT<sub>x</sub>-10.

<sup>3</sup> The units were tested at different times and the subscripts on the UUT's reference the following lab test reports:

x - 1340-05 - Sept 2022 / y - 1340-01 - Nov 2019

<b>SEISMIC CERTIFICATION LIMITS</b>								
System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>	F <sub>P</sub> / W <sub>P</sub>
CT Gantries	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Patient Tables	CBC 2022	2.0	1.0	1.50	1.0	1.5	1.5	2.40
Power Supply Cabinets	CBC 2022	2.0	1.0	1.50	2.5	6.0	2.0	1.50



<b>Table 1</b>	<b>UNITED IMAGING HEALTHCARE Co., Ltd</b>					 <b>WEGAI</b> <small>W.E. GUNDY &amp; ASSOCIATES, INC.</small> <small>STRUCTURAL &amp; EARTHQUAKE ENGINEERING</small>			
	<b>SPECIAL SEISMIC CERTIFICATION</b>								
<b>CERTIFIED SYSTEM AND COMPONENTS</b>									

<b>System: ATLAS uCT Systems</b>					<b>Manufacturer: United Imaging Healthcare Co., Ltd</b>				
----------------------------------	--	--	--	--	---	--	--	--	--

System Component <sup>1</sup>	United Imaging Part Number	Dimensions (in)			Weight (lb)	Mounting	UUT <sup>3</sup>
		Width	Length	Height			

<b>Reconstruction System</b>							
------------------------------	--	--	--	--	--	--	--

RSC - ATLAS	18023946	23.6	39.4	49.2	396	floor	UUT <sub>x</sub> -12
-------------	----------	------	------	------	-----	-------	----------------------

<b>Imaging Systems (Cabinet / Internally Mounted PC)<sup>4</sup></b>							
--	--	--	--	--	--	--	--

PC Cabinet PC2 - uCT / uMI	18022538 60004342	34.2 7.0	30.3 19.0	35.8 18.6	174 37	floor	UUT <sub>y</sub> -10
PC Cabinet PC1 - uCT	18022537 60003853	27.9 7.0	26.7 17.2	27.2 20.5	176 44	floor	UUT <sub>y</sub> -9
PC Cabinet PC1 - ATLAS	18022536 60009576	36.2 7.0	34.6 17.2	32.4 20.5	276 37	floor	interpolated
PC Cabinet PC3 - uCT / uMI	18022536 80504016	36.2 7.0	34.6 17.8	32.4 25.5	276 56	floor	UUT <sub>y</sub> -11

**Notes:**

<sup>1</sup> All components are manufactured by United Imaging Healthcare Co., Ltd. The part numbers listed uniquely identify the type of component, manufacturer, and material of construction for each sub-component within the tested units.

<sup>2</sup> Patient table weights do not include 362lb simulated patient weight for UUT<sub>y</sub>-5 and 565lbs for UUT<sub>x</sub>-10.

<sup>3</sup> The units were tested at different times and the subscripts on the UUT's reference the following lab test reports:  
x - 1340-05 - Sept 2022 / y - 1340-01-RI - Nov 2019

<sup>4</sup> PCs are installed in any of three cabinets (see model numbers below and dimensions / weights above) by United Imaging Healthcare.

18022536 / 18022537 / 18022538

<b>SEISMIC CERTIFICATION LIMITS</b>								
-------------------------------------	--	--	--	--	--	--	--	--

System Component	Code	S <sub>DS</sub> (g)	z / h	I <sub>P</sub>	a <sub>P</sub>	R <sub>P</sub>	Ω <sub>0</sub>	F <sub>P</sub> / W <sub>P</sub>
Reconstruction System	CBC 2022	2.0	1.0	1.50	2.5	6.0	2.0	1.50
PC / User Interface	CBC 2022	2.0	1.0	1.50	2.5	6.0	2.0	1.50

UUT<sub>x-9</sub>

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (8) M12 grade 8.8 bolts



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> uCT ATLAS	<b>Test Date:</b> September 2022
<b>Model Number:</b> 7808924	<b>Report Number:</b> 1340-05
<b>UUT Function:</b> Continuous rotating x-ray to generate diagnostic imaging	
<b>UUT Description:</b> Gantry for the ATLAS uCT systems.	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
5,732	91.8	36.6	77.8	8.5	14.5	28.5

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>Ds</sub> (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.

UUT<sub>y</sub>-2

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (4) M16 grade 12.9 bolts



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> uCT 760/780 CT Gantry	<b>Test Date:</b> November 2019
<b>Model Number:</b> 78001101	<b>Report Number:</b> 1340-01
<b>UUT Function:</b> Continuous rotating x-ray to generate diagnostic imaging	
<b>UUT Description:</b> Component of the uCT 760/780 CT system	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
5,157	31.5	91.7	77.8	9.5	6.0	23.5

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>ds</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.




UUT<sub>y</sub>-5

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (4) M12 grade 12.9 bolts



BY: Mohammad Karim

<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> uCT – PS2	<b>Test Date:</b> September 2019
<b>Model Number:</b> 80100042	<b>Report Number:</b> 1340-01
<b>UUT Function:</b> Motorized patient support	
<b>UUT Description:</b> Component of the uCT CT systems	

**UUT PROPERTIES**

Weight (lb) with Patient	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,422	23.6	161.6	37.4	8.0	9.0	8.5

The patient table moves vertically / horizontally to accommodate different positions and procedures. The system was tested in the normal operating position, the table extended 41.7", at the maximum height, and a total simulated patient weight of 362lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	SDS (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.

UUT<sub>x</sub>-10

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (12) M12 grade 8.8 bolts



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> uCT PS4	<b>Test Date:</b> September 2022
<b>Model Number:</b> 78008392	<b>Report Number:</b> 1340-05
<b>UUT Function:</b> Motorized patient support	
<b>UUT Description:</b> Patient support for the ATLAS uCT systems.	

**UUT PROPERTIES**

Weight (lb) with Patient	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
2,660	22.1	169.3	37.4	3.5	4.5	7.0

The patient table moves vertically / horizontally to accommodate different positions and procedures. The system was tested in the normal operating position, with the table extended 49.8", at the maximum height, and a total simulated patient weight of 565lbs.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>Ds</sub> (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67


Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.

UUT<sub>x</sub>-11

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (8) M12 grade 12.9 bolts thru United Imaging seismic anchorage brackets.



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> PCS - ATLAS	<b>Test Date:</b> September 2022
<b>Model Number:</b> 78004466	<b>Report Number:</b> 1340-05
<b>UUT Function:</b> Power Supply Cabinet	
<b>UUT Description:</b> Power supply for the ATLAS uCT systems.	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
1,390	27.6	29.6	59.5	15.0	15.0	18.5

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>Ds</sub> (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.



UUT<sub>x</sub>-12

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Rigid floor mounted with (4) M12 grade 12.9 bolts thru (2) 3mm thick United Imaging U brackets mounted to the cabinet with (2) M12 screws per bracket.



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> RSC - ATLAS	<b>Test Date:</b> September 2022
<b>Model Number:</b> 18023946	<b>Report Number:</b> 1340-05

**UUT Function:** Image reconstruction system

**UUT Description:** Image reconstruction system cabinet for the ATLAS uCT systems. The cabinet housed two PCs used for image reconstruction with model numbers 60005332 and 60009013.

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
396	23.6	39.4	49.2	15.0	6.5	23.0

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>ds</sub> (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.

UUT<sub>y</sub>-9

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Cabinet is Rigid floor mounted with (4) M12 grade 12.9 bolts thru United Imaging seismic anchorage brackets (18042254). Internal PC mounts to the cabinet with United Imaging seismic restraint brackets (18042256).



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> PC Cabinet / Internal PC1 – uCT	<b>Test Date:</b> November 2019
<b>Model Number:</b> 18022537 / 60003853	<b>Report Number:</b> 1340-01
<b>UUT Function:</b> Imaging System PC	
<b>UUT Description:</b> Component of the uCT CT systems.	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
176	27.9	26.7	27.2	>33.0	16.0	22.5
44	7.0	20.5	17.2	NA	NA	NA

Note: The first line above details the UUT Properties for the PC Cabinet and the second line details the UUT properties for the internally mounted PC.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	SDS (g)	z / h	I <sub>p</sub>	AFLX-H (g)	ARIG-H (g)	AFLX-V (g)	ARIG-V (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.

UUT<sub>y</sub>-10

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Cabinet is Rigid floor mounted with (4) M12 grade 12.9 bolts thru United Imaging seismic anchorage brackets (18042255). Internal PC mounts to the cabinet with United Imaging seismic restraint brackets (18042256).



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> PC Cabinet / PC2 – uCT / uMI	<b>Test Date:</b> November 2019
<b>Model Number:</b> 18022538 / 60004342	<b>Report Number:</b> 1340-01
<b>UUT Function:</b> Imaging System PC	
<b>UUT Description:</b> Component of the uCT and uMI CT systems.	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
174	34.2	30.3	35.8	>33.0	12.0	>33.0
37	7.0	19.0	18.6	NA	NA	NA

Note: The first line above details the UUT Properties for the PC Cabinet and the second line details the UUT properties for the internally mounted PC.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.




UUT<sub>y</sub>-11

**UNIT UNDER TEST (UUT)  
SUMMARY SHEET**



**Mounting Details:** Cabinet is Rigid floor mounted with (4) M12 grade 12.9 bolts thru United Imaging seismic anchorage brackets (18042255). Internal PC mounts to the cabinet with United Imaging seismic restraint brackets (18042256).



<b>Manufacturer:</b> United Imaging Healthcare	<b>Test Location:</b> EETI (Suzhou, CHINA)
<b>Component:</b> PC Cabinet / PC3 – uCT / uMI	<b>Test Date:</b> November 2019
<b>Model Number:</b> 18022536 / 80504016	<b>Report Number:</b> 1340-01
<b>UUT Function:</b> Imaging System PC	
<b>UUT Description:</b> Component of the uCT and uMI CT systems.	

**UUT PROPERTIES**

Weight (lb)	Dimensions (inches)			Natural Frequency (Hz)		
	Width	Depth	Height	FB	SS	V
276	36.2	34.6	32.4	13.0	9.5	17.0
56	7.0	17.8	25.5	NA	NA	NA

Note: The first line above details the UUT Properties for the PC Cabinet and the second line details the UUT properties for the internally mounted PC.

**SEISMIC TEST PARAMETERS**

Building Code / Test Criteria	S <sub>DS</sub> (g)	z / h	I <sub>p</sub>	A <sub>FLX-H</sub> (g)	A <sub>RIG-H</sub> (g)	A <sub>FLX-V</sub> (g)	A <sub>RIG-V</sub> (g)
CBC 2022 / ICC-ES AC156	2.00	1.0	1.5	3.20	2.40	-	-
	2.50	0	1.5	-	-	1.67	0.67

Note: The unit was full of contents during testing and remained functional before and after the ICC-ES AC156 test. The unit maintained structural integrity during and after the ICC-ES AC156 test.