



DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT

**APPLICATION FOR HCAI SPECIAL SEISMIC  
CERTIFICATION PREAPPROVAL (OSP)**

OFFICE USE ONLY

APPLICATION #: OSP-0878

**HCAI Special Seismic Certification Preapproval (OSP)**

Type:  New  Renewal

**Manufacturer Information**

Manufacturer: GE Healthcare

Manufacturer's Technical Representative: Tom Farnow

Mailing Address: 3000 N. Grandview Blvd., Waukesha, WY 53188

Telephone: (888) 406-1101 Email: tom.farnow@gehcseismic.com

**Product Information**

Product Name: Aurora

Product Model Number(s): See Attachment 1

Product Category: CT Systems

Product Sub-Category: CT Systems

General Description: System components of a multiple-component SPECT/CT imaging system. Special Seismic Certification is limited to components of the system identified in Attachment 1 for functions related to Computed Tomography (CT) imaging services for the diagnostic assessment of trauma.

Mounting Description: Base Mounted Rigid

Tested Seismic Enhancements: Seismic enhancements made to the test units and/or modifications required to address anomalies during the tests shall be incorporated into the production units.

**Applicant Information**

Applicant Company Name: EASE LLC

Contact Person: Jonathan Roberson

Mailing Address: 5877 Pine Ave. Suite 210, Chino Hills, CA 91709

Telephone: (909) 606-7622 Email: jon@easeco.com

Title: Principal Structural Engineer



**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT**

**California Licensed Structural Engineer Responsible for the Engineering and Test Report(s)**

Company Name: EASE LLC

Name: Jonathan Roberson

California License Number: S4197

Mailing Address: 5877 Pine Ave., Suite 210, Chino Hills, CA 91709

Telephone: (951) 295-1892

Email: jon@EASECo.com

**Certification Method**

GR-63-Core

ICC-ES AC156

IEEE 344

IEEE 693

NEBS 3

Other (Please Specify): \_\_\_\_\_

**Testing Laboratory**

Company Name: ENVIRONMENTAL TESTING LABORATORIES, INC. (ETL)

Contact Person: Jeremy Lange


Mailing Address: 11034 Indian Trail, Dallas TX 75229-3513

Telephone: (972) 247-9657

Email: jeremy@etldallas.com

BY: Mohammad Karim

DATE: 04/10/2026





**DEPARTMENT OF HEALTH CARE ACCESS AND INFORMATION  
OFFICE OF STATEWIDE HOSPITAL PLANNING AND DEVELOPMENT**

**Seismic Parameters**

Certified Response Spectral Acceleration Factors:(F<sub>p</sub>/W<sub>p</sub>)

Horizontal	(A Flx-H), g=	<u>3.20</u>	(A Rig-H), g=	<u>2.15</u>
Vertical	(A Flx-V), g=	<u>1.68</u>	(A Rig-V), g=	<u>0.68</u>

SDS (Design spectral response acceleration at short period, g) = 2.5 @ z/h=0; 2.00 @ z/h=1

H<sub>f</sub> (Force amplification height factor) = 1 @ z/h=0; 3.5 @ z/h=1

R<sub>u</sub> (Structure ductility reduction factor) = 1 @ z/h=0; 1.3 @ z/h=1

I<sub>p</sub> (Importance factor) = 1.5

z/h (Height ratio factor) = 0 and 1

**HCAI Approval (For Office Use Only) - Approval Expires on 04/12/2032**


Date: 4/10/2026

Name: Mohammad Karim Title: Supervisor, Health Facilities

Condition of Approval (if applicable): OSP-0878

BY: Mohammad Karim

DATE: 04/10/2026



**ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS**

ATTACHMENT PAGE | 1 OF 2

**TABLE 1:**

**BUILDING CODE: 2025 CBC**

Manufacturer		GE Healthcare			Seismic Parameters			
System		Aurora			S <sub>DS</sub> = 2.00 at z/h = 1		R <sub>μ</sub> = 1.3; H <sub>f</sub> = 3.5	
Product Type		NM/CT System			S <sub>DS</sub> = 2.50 at z/h = 0		R <sub>μ</sub> = 1.0; H <sub>f</sub> = 1.0	
COMPONENT	MODEL NO.	DIMENSIONS (IN.)			APPROX. WT. (LB.)	MOUNT	BASIS <sup>[1]</sup>	
		W	D	H				
<b>GANTRIES</b>								
Aurora NM Gantry	5376204-70-80 <sup>[2]</sup>	84.7	120.2	78.7	8808	Rigid Base	UUT-2517-1	
Aurora Revolution Ascend CT Gantry	6966000-350 <sup>[3]</sup>							
Aurora NM Gantry	5376204-70-80 <sup>[2]</sup>	84.7	120.2	78.7	8808	Rigid Base	SAME	
Aurora Revolution Ascend CT Gantry	6966000-150 <sup>[3]</sup>							
<b>PATIENT TABLES</b>								
Aurora Hybrid Patient Table	5363260-80 <sup>[4]</sup>	24	110.1	44.4	1228	Rigid Base	UUT-2517-2	
<b>CONSOLES</b>								
Z4G5 NM Console (HP)	5966435-11 <sup>[5]</sup>	6.65	17.5	15.2	26	Rigid Base	UUT-2517-3	
NM Operator Console (HP)	5370540-95 <sup>[6]</sup>	6.65	17.5	15.2	25	Rigid Base	UUT-1819-8	
Z4G4 Aurora NM Console (HP)	5966435-01 <sup>[5]</sup>	6.65	17.5	15.2	25	Rigid Base	SAME	
Z4G4 Aurora NM Console (HP)	5966435-02 <sup>[5]</sup>	6.65	17.5	15.2	25	Rigid Base	SAME	
Operator Console – Standalone (HP)	5940104-31 <sup>[7]</sup>	23.6	25.6	21.0	132	Rigid Base	UUT-2201-2	
Operator Console – Standalone (HP)	5940104-11 <sup>[8]</sup>	23.6	25.6	21.0	132	Rigid Base	SAME	
Operator Console – Standalone (HP)	5940104-21 <sup>[8]</sup>	23.6	25.6	21.0	132	Rigid Base	SAME	
Operator Console – Standalone II (HP)	5940104-12 <sup>[8]</sup>	23.6	25.6	21.0	130	Rigid Base	UUT-2416-1	
Operator Console – Standalone II (HP)	5940104-22 <sup>[8]</sup>	23.6	25.6	21.0	130	Rigid Base	SAME	
<b>POWER DISTRIBUTION</b>								
Power Distribution Unit (PDU)	2326492-80	27.6	21.7	41.8	841 <sup>[10]</sup>	Rigid Base	UUT-1126-3	
Power Distribution Unit (PDU)	2326492-81	27.6	21.7	41.8	841	Rigid Base	SAME	
Power Input Distribution Box (PIDB)	5375888-01	19.7	6.13	19.7	22.5	Wall	UUT-1133-3	
Power Input Distribution Box (PIDB)	5375888	19.7	6.13	19.7	22.5	Wall	SAME	

Table continues next page.

**ATTACHMENT 1: SEISMIC CERTIFIED COMPONENTS**


**TABLE 1:** (continued)


**BUILDING CODE: 2025 CBC**

<i>Manufacturer</i>	<b>GE Healthcare</b>				<b>Seismic Parameters</b>		
<i>System</i>	<b>Aurora</b>				S <sub>DS</sub> = 2.00 at z/h = 1      R <sub>μ</sub> = 1.3; H <sub>f</sub> = 3.5		
<i>Product Type</i>	<b>NM/CT System</b>				S <sub>DS</sub> = 2.50 at z/h = 0      R <sub>μ</sub> = 1.0; H <sub>f</sub> = 1.0		
COMPONENT	MODEL NO.	DIMENSIONS (IN.)			APPROX. WT. (LB.)	MOUNT	BASIS <sup>[1]</sup>
		W	D	H			
<i>Mount</i>	<p><u>Rigid Base</u>: free-standing, base-mounted tower configuration with the component rigidly attached to a supporting structure and no lateral support above the base.</p> <p><u>Wall</u>: fully supported by a rigid building wall or partition.</p>						
<i>Notes</i>	<ol style="list-style-type: none"> <li>BASIS: <ul style="list-style-type: none"> <li>UUT-#: Indicates that a test specimen matching these characteristics was tested. See Attachment 2.</li> <li>SAME: Indicates model is physically, mechanically &amp; electrically the same as test specimen. Differences are limited to color, software, and/or identification number.</li> </ul> </li> <li>Requires use of seismic kits: P/N 6001790 &amp; P/N 5824496.</li> <li>Requires use of seismic kit: P/N 6001676.</li> <li>Requires use of seismic kit: P/N 6001675.</li> <li>Requires use of seismic kit: P/N 5973645.</li> <li><i>Requires use of GE Healthcare System Seismic Kit H3909CY or equivalent (NM/CT850/860).</i></li> <li><i>Requires use of seismic kit 5881922 (Commercial Catalog B77162RE).</i></li> <li>Requires use of seismic kit: P/N 5873550.</li> <li>All seismic kits identified above for components that are part of the Aurora System are included in the collector seismic kit: Aurora System Seismic Option P/N 6001674 (also referenced as GE Healthcare Catalog No. H3913DD &amp; B84792DA).</li> <li>Weight includes mounting brackets used in test.</li> <li><i>Gray italic formatting indicates unit is not a component of system and used as basis for seismic qualification only.</i></li> </ol>						


**ATTACHMENT 2: TEST SPECIMEN SUMMARY**


ATTACHMENT PAGE | 1 OF 4

<b>UUT-2517-1 Aurora NM/CT Gantry</b>									
MANUFACTURER: GE Medical Systems									
IDENTIFICATION:									
		Model No.	Serial No.						
NM Gantry		5376204-70-80	830D60703						
Revolution Ascend Plus CT Gantry		6966000-350	NGXAH2500001CN						
DESCRIPTION:									
Component of Aurora NM/CT System									
NM Gantry is production equivalent engineering unit CT Gantry is standard production  NM Gantry requires use of seismic kits: P/N 6001790 & P/N 5824496. CT Gantry requires use of seismic kit: P/N 6001676. *Measured weight is within 0.5% of Reported Wt.= 8808 lb.									
MOUNTING:									
Rigid Base (Floor) mounted using (8) – M10 Class 8.8 bolts through steel sleeves extracted from Hilti HSL4 (or HSL3) anchors. Torque = 18 ft-lb. (CT Gantry) and (4) – M12 Class 8.8 bolts through steel sleeves extracted from Hilti HSL4 (or HSL3) anchors. Torque = 44 ft-lb. (NM Gantry)									
Dimensions (in.)					Lowest Resonant Frequency (Hz)				
Width	Depth	Height	Weight (lb)		Side (X) - Axis	Front (Y) - Axis	Vertical (Z) - Axis		
84.7	120.2	78.7	8765*		4.36	4.75	4.81		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters								Building Code: 2025 CBC	
S <sub>Ds</sub>	z/h	R <sub>μ</sub>	H <sub>r</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00	1	1.3	3.5	1.5	3.20	2.40	1.68	0.68	
2.50	0	1.0	1.0						
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									


<b>UUT-2517-2 Aurora Hybrid Patient Table</b>									
MANUFACTURER: GE Medical Systems Israel									
IDENTIFICATION:									
		Model No.:	Serial No.:						
		5363260-80	MAVSHO001						
DESCRIPTION:									
Component of Aurora NM/CT System									
Requires use of seismic kit: P/N 6001675.									
MOUNTING:									
Rigid Base (Floor) mounted using (20) – M10 Class 8.8 Bolts in steel sleeve extracted from Hilti HSL-3 M10/20 anchors									
Dimensions (in.)					Lowest Resonant Frequency (Hz)				
Width	Depth	Height	Weight (lb)		Side (X) - Axis	Front (Y) - Axis	Vertical (Z) - Axis		
24.0	110.1 – 246.3*	44.4	1229 †		2.97	24.72	3.17		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters								Building Code: 2025 CBC	
S <sub>Ds</sub>	z/h	R <sub>μ</sub>	H <sub>r</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00	1	1.3	3.5	1.5	3.20	2.40	1.68	0.68	
2.50	0	1.0	1.0						
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									


**ATTACHMENT 2: TEST SPECIMEN SUMMARY**

<b>UUT-2517-3 NM Console (Z4G5)</b>									
MANUFACTURER: Hewlett Packard									
IDENTIFICATION: GE Label Model No.: 5966435-11 Serial No.: CZC5217TFK HP Label Product No.: 57K35AV Serial No.: CZC5217TFK									
DESCRIPTION: Component of Aurora NM/CT System.  Requires use of seismic kit: P/N 5973645.									
MOUNTING: Rigid Base (Floor) mounted using a strap assembly at front and rear. Each strap assembly consisting of: <ul style="list-style-type: none"> <li>1" wide polyester strap w/ metal cam buckle (200 lb working load) looped through slots in vertical legs of</li> <li>(2) – L2½ x 2" x 4.5mm x 0'-2" LLV brackets positioned snug against unit.</li> </ul> Each bracket attached w/ 3/8" diameter SAE J429 Grade 8 Hex Head bolt w/ Standard Washer threaded into aluminum plate. (1 set ea. front & rear, 4 bolts total.)									
Dimensions (in.)				Weight (lb)	Lowest Resonant Frequency (Hz)				
Width	Depth	Height			Front (X) - Axis	Side (Y) - Axis	Vertical (Z) - Axis		
6.65	17.5	15.2		24.0 w/o mount kit 26.1 w/ mount kit	32.70	17.41	12.10		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters								Building Code: 2025 CBC	
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>f</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00 2.50	1 0	1.3 1.0	3.5 1.0	1.5	3.20	2.40	1.68	0.68	
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									


<b>UUT-1819-8 NM Operator Console (HP)</b>									
MANUFACTURER: Hewlett Packard									
IDENTIFICATION: Model No.: 5370540-95 Serial No.: CZC52800CT									
DESCRIPTION: Component of the NM/CT 860 system; included as basis for component of Aurora NM/CT System.  Component used in several GE Healthcare NM/CT systems: NM/CT 870 DR, NM/CT 870 CZT, NM/CT 860, NM/CT 850 HP Z4 G4 computer.									
MOUNTING: Rigid Base (Floor) mounted using a strap assembly at front and rear. Each strap assembly consisting of: <ul style="list-style-type: none"> <li>1" wide nylon strap w/ metal cam buckle (50 lb tension / 200 lb WLL) looped through slots in vertical legs of</li> <li>(2) – L2½ x 2" x 1/8" x 0'-2" LLV brackets positioned snug against unit.</li> </ul> Each bracket attached w/ 3/8" diameter SAE J429 Grade 8 threaded into aluminum plate. (1 set ea. front & rear, 4 bolts total.)									
Dimensions (in.)				Weight (lb)	Lowest Resonant Frequency (Hz)				
Width	Depth	Height			Front (X) - Axis	Side (Y) - Axis	Vertical (Z) - Axis		
6.65	17.5	15.2		25	32.89	30.96	35.15		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters								Building Code: 2025 CBC	
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>f</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00 2.60	1 0	1.3 1.0	3.5 1.0	1.5	3.20	2.40	1.74	0.70	
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									


**ATTACHMENT 2: TEST SPECIMEN SUMMARY**

<b>UUT-2201-2 Operator Console - Standalone</b>									
<b>MANUFACTURER:</b> GE Healthcare Japan Corporation									
<b>IDENTIFICATION:</b> Model No.: 5940104-31 Serial No.: SACENG002									
<b>DESCRIPTION:</b> Component of Revolution Ascend CT System; included as basis for component of Aurora NM/CT System.  <u>Includes:</u> HP Z8 G4 Workstation (ID No.: 5809620-7 Serial No.: 4CV907WMT2) GE Healthcare Power Box (ID No.: 5808401-2 Serial No.: DW2003029)  Unit requires the use of Seismic Kit 5881922 which contains mounting bracket hardware.									
<b>MOUNTING:</b> <u>Rigid Base:</u> mounted using (3) – 3/8" dia. hex head bolts to interface plate.									
Dimensions (in.)				Weight (lb)	Lowest Resonant Frequency (Hz)				
Width	Depth	Height			Side (X) - Axis	Front (Y) - Axis	Vertical (Z) - Axis		
23.6	25.6	21.0		132	24.62	31.57	20.89		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters							Building Code: 2025 CBC		
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>r</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00	1	1.3	3.5	1.5	3.20	2.40	1.68	0.68	
2.50	0	1.0	1.0						
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									

<b>UUT-2416-1 Operator Console – Standalone II</b>									
<b>MANUFACTURER:</b> GE Hangwei Medical Systems Co., LTD									
<b>IDENTIFICATION:</b> Model No.: 5940104-12 Serial No.: 7DC497S369									
<b>DESCRIPTION:</b> Component of Revolution Ascend CT Family (as tested).  <u>Includes:</u> HP Z8 G5 Workstation (ID No.: 67V63AV Serial No.: 4CE339CG3Z) GE Healthcare Power Box (ID No.: 5808401-2 Serial No.: DW2402030)  Unit requires the use of Seismic Kit 5873550 which contains mounting bracket hardware.									
<b>MOUNTING:</b> <u>Rigid Base:</u> mounted using (3) – 3/8" dia. hex head bolts to interface plate.									
Dimensions (in.)				Weight (lb)	Lowest Resonant Frequency (Hz)				
Width	Depth	Height			Side (X) - Axis	Front (Y) - Axis	Vertical (Z) - Axis		
23.6	25.6	20.9		129.5	10.4	17.9	14.1		
ICC-ES AC156 (24) 2 <sup>nd</sup> Edition Test Parameters							Building Code: 2025 CBC		
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>r</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00	1	1.3	3.5	1.5	3.20	2.40	1.68	0.68	
2.50	0	1.0	1.0						
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									

**ATTACHMENT 2: TEST SPECIMEN SUMMARY**

<b>UUT-1126-3 Power Distribution Unit (PDU)</b>									
<b>MANUFACTURER:</b> GE Hangwei Medical Systems Co., LTD									
<b>IDENTIFICATION:</b> Model No.: 2326492-80 Serial No.: 270337HM9									
<b>DESCRIPTION:</b> Component of the Optima CT660 System; included as basis for component of Aurora NM/CT System.									
<b>MOUNTING:</b> Rigid Base (Floor) mounted using: (2) – 3/8" dia. ASTM A574 Socket Head Cap Screws w/ washer through each GE mounting bracket to floor plate. (4 anchors total) GE mounting assembly including: (2) – Seismic Bracket, System Cabinet (P/N 2354563-2) Each bracket mounted to cabinet w/: (2) – M10 x 25mm Class 12.9 bolt (Torque= 38.4 N-m) (2) – M10 Lock Washer (P/N 2203-M10-07) (2) – M10 Flat Washer (P/N 2000-M10-03)									
<b>Dimensions (in.)</b>				<b>Lowest Resonant Frequency (Hz)</b>					
Width	Depth	Height	Weight (lb)	Front (X) - Axis	Side (Y) - Axis	Vertical (Z) - Axis			
27.6	21.7	41.8	841	20.4	20.6	13.8			
<b>ICC-ES AC156 (24) 2<sup>nd</sup> Edition Test Parameters</b>								<b>Building Code: 2025 CBC</b>	
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>f</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00 2.50	1 0	1.3 1.0	3.5 1.0	1.5	3.20	2.40	1.68	0.68	
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									

<b>UUT-1133-3 Power Input Distribution Box (PIDB)</b>									
<b>MANUFACTURER:</b> GE Medical Systems									
<b>IDENTIFICATION:</b> Model No.: 5375888-01 Serial No.: WC0180225/10									
<b>DESCRIPTION:</b> Component of the Discovery NM-CT 670 System; included as basis for component of Aurora NM/CT System.									
<b>MOUNTING:</b> Wall Mounted w/ (4) - #10 Tek screws w/ 1.25" O.D. fender washers through 5/8" gypsum board into 16 ga. steel stud.									
<b>Dimensions (in.)</b>				<b>Lowest Resonant Frequency (Hz)</b>					
Width	Depth	Height	Weight (lb)	Side (X) - Axis	Front (Y) - Axis	Vertical (Z) - Axis			
19.69	6.13	19.69	22.5	N/A	N/A	N/A			
<b>ICC-ES AC156 (24) 2<sup>nd</sup> Edition Test Parameters</b>								<b>Building Code: 2025 CBC</b>	
S <sub>DS</sub>	z/h	R <sub>μ</sub>	H <sub>f</sub>	I <sub>p</sub>	A <sub>FLX-H</sub>	A <sub>RIG-H</sub>	A <sub>FLX-V</sub>	A <sub>RIG-V</sub>	
2.00 2.60	1 0	1.3 1.0	3.5 1.0	1.5	3.20	2.40	1.74	0.70	
Unit satisfied AC156 requirements for structural integrity and manufacturer requirements for functionality after AC156 test.									

**ATTACHMENT 3: UUT SUMMARY INDEX**

ATTACHMENT PAGE | 1 OF 1

UUT	REPORT NUMBER	TEST DATE	TEST FACILITY	REPORT UUT
2517-1	SQ12251701	October 30 - 31, 2025	Environmental Testing Laboratory, Inc.	1
2517-2				2
2517-3				3
1819-8	SQ12181901	March 26, 2019	Environmental Testing Laboratory, Inc.	8
2201-2	SQ12220101	February 21 - 23, 2022	Environmental Testing Laboratory, Inc.	2
2416-1	SQ12241601	August 26, 2024	Environmental Testing Laboratory, Inc.	1
1126-3	SQ12112601	October 19 - 20, 2011	Environmental Testing Laboratory, Inc.	3
1133-3	SQ12113301	Nov 30 - Dec 2, 2011	Environmental Testing Laboratory, Inc.	3

